

## Technical Publications

**Direction: 5335141** 

Revision: 01

## LOGIQ P6

# Integrating the Healthcare Enterprise (IHE) Integration Statement

#### INTRODUCTION

#### **OVERVIEW**

This IHE Integration Statement describes the intended conformance of this GE product with the IHE Technical Framework. The IHE Technical Framework defines a subset of the functional components of the healthcare enterprise, called IHE Actors, and specifies their interactions in terms of a set of coordinated, standards-based transactions. These transactions are organized into functional units called Integration Profiles that highlight their capacity to address specific clinical needs.

Users familiar with these concepts can use Integration Statements as an aid to determine what level of integration this product supports with other systems and what clinical and operational benefits such integration might provide.

Other GE Integration Statements as well as further information on IHE may be found on Internet at: **ge.com/ihe** 

#### IMPORTANT REMARKS

The use of these IHE Integration Statements, by itself, it is not sufficient to ensure that inter-operation will be successful. The user needs to proceed with caution and ensure that a qualified systems' integration (such as that provided by GE's IT Professional Services) address at least four issues:

- Integration The integration of any system into an integrated department or enterprise requires an analysis of the applications requirements. The design of a solution that integrates GE systems with non–GE systems is the user's responsibility and should not be underestimated. Special care should be exercised in defining the expected workflow, the special conditions, and the clinical practice to be supported to ensure a safe and effective operation.
- Validation Testing the complete range of possible interactions between any GE system and non-GE system, before the integration is declared operational, should not be overlooked. Therefore, the user should ensure that any non-GE provider accepts full responsibility for all validation required for their connection with GE systems. This includes the accuracy of the data once it has crossed the interface and the stability of the data for the intended applications. Such a validation is required before any clinical use (diagnosis and/or treatment) is performed.
- Future Evolution GE understands that the IHE Technical Framework and the underlying communication standards will evolve to meet the user's growing requirements. GE is actively involved in the development of the IHE Technical Framework and the underlying standards. Such evolutions may require changes to already installed systems. In addition, GE reserves the right to discontinue or make changes to product features including those reflected by this IHE Integration Statement. The user should ensure that any non-GE system that connects with GE systems also plans for the future evolution of IHE. Failure to do so will likely result in the loss of function and/or connectivity as the IHE Integration profiles evolve and GE Products are enhanced to support these changes.

### **IHE INTEGRATION STATEMENT**

Genera	Date					
	March25,2008					
Vendo	r	Product Name		ersion		
GE Health	care	LOGIQ P6	R1.0.X			
This product impleme Profiles, Actors and C		s required in the IHE Technical Framewow:	work to support the I	HE Integration		
Integration Profiles Actors Implemented Implemented		Options	Options Implemented			
Scheduled Workflow		Acquisition Modality	Broad Wor	Broad Worklist Query		
			Patient Bas Query	Patient Based Worklist Query		
		Image Display	none			
PIR		Acquisition Modality	None			
ED		Evidence Creator	None			
		Image Display	None			
Internet address for	or GE Healthca	re IHE information: http://www.geh	ealthcare.com/usen/i	nteroperability/		
Lin	ks to Standard	s Conformance Statements for th	e Implementation			
HL7	Not Applicable					
DICOM	http://www.gehealthcare.com/usen/interoperability/dicom/docs/5308877_100r1.pdf					
	L	inks to general information on IH	E			
In North America: www.ihe.net www.himss.org/ihe		In Europe: www.ihe-europe.org	In Japan: www.jira	-net.or.jp/ihe-j		

This page is left blank



#### **GE Healthcare**

GE Healthcare – Americas : Fax 262-544-3384 P.O.Box 414, Milwaukee, Wisconsin, 53201 U.S.A Internet – http://www.gehealthcare.com

GE Healthcare – Europe : Fax 33-1-30-70-95-95 Paris, France Internet - http://www.gehealthcareeurope.com

GE Healthcare – Asia Tokyo, Japan : Fax 81-3-3223-8560 Singapore : Fax 65-291-7006