



# Technical Publications

Document Number: DOC1891675

Revision: 1

DoseWatch

Version 2.1

Integrating the Healthcare Enterprise (IHE) Integration Statement

### INTRODUCTION

#### **OVERVIEW**

This IHE Integration Statement describes the intended conformance of this GE product with the IHE Technical Framework. The IHE Technical Framework defines a subset of the functional components of the healthcare enterprise, called IHE Actors, and specifies their interactions in terms of a set of coordinated, standards-based transactions. These transactions are organized into functional units called Integration Profiles that highlight their capacity to address specific clinical needs.

Users familiar with these concepts can use Integration Statements as an aid to determine what level of integration this product supports with other systems and what clinical and operational benefits such integration might provide.

Other GE Integration Statements as well as further information on IHE may be found on Internet at: <a href="http://www.ge.com/ihe">http://www.ge.com/ihe</a>

#### IMPORTANT REMARKS

The use of these IHE Integration Statements, by itself, it is not sufficient to ensure that inter-operation will be successful. The user needs to proceed with caution and ensure that a qualified systems' integration (such as that provided by GE's IT Professional Services) address at least three issues:

- Integration The integration of any system into an integrated department or enterprise requires an analysis of the applications requirements. The design of a solution that integrates GE systems with non-GE systems is the user's responsibility and should not be underestimated. Special care should be exercised in defining the expected workflow, the special conditions, and the clinical practice to be supported to ensure a safe and effective operation.
- Validation Testing the complete range of possible interactions between any GE system and non-GE system, before the integration is declared operational, should not be overlooked. Therefore, the user should ensure that any non-GE provider accepts full responsibility for all validation required for their connection with GE systems. This includes the accuracy of the data once it has crossed the interface and the stability of the data for the intended applications. Such a validation is required before any clinical use (diagnosis and/or treatment) is performed.
- Future Evolution GE understands that the IHE Technical Framework and the underlying communication standards will evolve to meet the user's growing requirements. GE is actively involved in the development of the IHE Technical Framework and the underlying standards. Such evolutions may require changes to already installed systems. In addition, GE reserves the right to discontinue or make changes to product features including those reflected by this IHE Integration Statement. The user should ensure that any non-GE system that connects with GE systems also plans for the future evolution of IHE. Failure to do so will likely result in the loss of function and/or connectivity as the IHE Integration profiles evolve and GE Products are enhanced to support these changes.

# **IHE INTEGRATION STATEMENT**

Gener	Date						
		c - IHE Integration Statemen		15 Apr 2016			
Vendor		Product Name	Version				
GE Healthcare		DoseWatch	V2.1				
This product implem Integration Profiles, A		tions required in the IHE Technical Framework tons listed below:	to support th	e IHE			
Integration Profiles Implemented		Actors Implemented	Options Implemented				
Access to Radiology Information		Image Display	None				
Consistent Time		Time Client	None				
Patient Administration Management		Patient Demographics Consumer	Merge				
Patient Identifier Cross- referencing for MPI		Patient Identifier Cross-reference Consumer	PIX Update Notification				
Patient Identifier Cross-reference for Mobile		Patient Identifier Cross-reference Consumer	None				
Radiation Exposure Monitoring		Dose Information Consumer	None				
		Dose Information Reporter	None				
		Dose Register	None				
Scheduled Workflow.b		Image Display	None				
		Performed Procedure Step Manager	None				
Internet address for	GE Healthcare	e IHE information: http://www.ge.com/ihe					
Links to Standards Conformance Statements for the Implementation							
DICOM	http://www3.gehealthcare.com/en/Products/ Interoperability/DICOM/Dose_Management_DICOM_Conformance_Statements						
Links to general information on IHE							
	IHE International: <u>www.ihe.net</u>						

This page is left blank

## Manufacturer:

GE Medical Systems - SCS
(A General Electric Company, going to market as GE Healthcare)
8, place de l'Hôpital
67000 Strasbourg
FRANCE

+33 3 88 22 80 40

